



Ref. Certif. No

UA-0194

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Capacitor charging power supply

Name and address of the applicant

OOO "OEM Tech", Odoevskogo Str.129, 220018
Minsk, Belarus

Name and address of the manufacturer

OOO "OEM Tech", Odoevskogo Str.129, 220018
Minsk, Belarus

Name and address of the factory

OOO "OEM Tech", Odoevskogo Str.129, 220018
Minsk, Belarus

Additional Information on page 2

Note: When more than one factory, please report on page 2

Ratings and principal characteristics

Input: 110-240VAC, 50-60Hz, 20.5-9.5A, Class I
Output: 500VDC

Trademark / Brand (if any)

OEM Tech

Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

PCA-20-500V-PD

Additional information (if necessary may also be reported on page 2)

The Risk Management requirements were not addressed. The usability evaluation has not been addressed
 Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005/AMD1:2012 inclusive National differences for USA, Canada

As shown in the Test Report Ref. No. which forms part of this Certificate

0492-2-2021

This CB Test Certificate is issued by the National Certification Body

SE "Ukrmetrteststandart"

4, Metrologichna Str.,
03143, Kiev, Ukraine

Date: 2021-07-06

Signature:

Victor Rymer



Ref. Certif. No

UA-0195

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
IECEE CB SCHEME

CB TEST CERTIFICATE

Product

Capacitor charging power supply

Name and address of the applicant

OOO "OEM Tech", Odoevskogo Str.129, 220018
Minsk, Belarus

Name and address of the manufacturer

OOO "OEM Tech", Odoevskogo Str.129, 220018
Minsk, Belarus

Name and address of the factory

OOO "OEM Tech", Odoevskogo Str.129, 220018
Minsk, Belarus

Additional information on page 2

Note: When more than one factory, please report on page 2

Ratings and principal characteristics

Input: 110-240VAC, 50-60Hz, 20.5-9.5A, Class I
Output: 800VDC

Trademark / Brand (if any)

OEM Tech

Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

PCA-20-800V-PD

Additional information (if necessary may also be reported on page 2)

The Risk Management requirements were not addressed. The usability evaluation has not been addressed
 Additional information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005/AMD1:2012 inclusive National differences for USA, Canada

As shown in the Test Report Ref. No. which forms part of this Certificate

0494-2-2021

This CB Test Certificate is issued by the National Certification Body

SE "Ukrmetrteststandart"
4, Metrologichna Str.,
03143, Kiev, Ukraine

Date: 2021-07-06

Signature:

Victor Rymer



Ref. Certif. No

UA-0196

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Capacitor charging power supply

Name and address of the applicant

OOO "OEM Tech", Odoevskogo 129-106, 220018
Minsk, Belarus

Name and address of the manufacturer

OOO "OEM Tech", Odoevskogo 129-106, 220018
Minsk, Belarus

Name and address of the factory

OOO "OEM Tech", Odoevskogo 129-106, 220018
Minsk, Belarus

Note: When more than one factory, please report on page 2

Additional Information on page 2

Ratings and principal characteristics

110-240VAC, 50-60Hz, 20,5-9,5 A, Class I
(Power output refer to the test report)

Trademark / Brand (if any)

OEM Tech

Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

PCA-20-500V-PD, PCA-20-800V-PD

Additional information (if necessary may also be reported on page 2)

Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1-2: 2014 inclusive National differences for
Canada

As shown in the Test Report Ref. No. which forms part of this Certificate

0491-5-2021

This CB Test Certificate is issued by the National Certification Body

SE "Ukrmetrteststandart"
4, Metrologichna Str.,
03143, Kiev, Ukraine

Date: 2021-07-14

Signature:

Victor Rymer


IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Capacitor charging power supply
Name and address of the applicant	OOO "OEM Tech" Odoevskogo 129-106, 220018 Minsk, Belarus
Name and address of the manufacturer	OOO "OEM Tech" Odoevskogo 129-106, 220018 Minsk, Belarus
Name and address of the factory	<input type="checkbox"/> Additional information on page 2 OOO "OEM Tech" Odoevskogo 129-106, 220018 Minsk, Belarus
Note: When more than one factory, please report on page 2	
Ratings and principal characteristics	Input: 110-240VAC, 50-60Hz, 20,5-9,5 A, Class I Output: 1000VDC
Trademark (if any)	OEM Tech
Customer's Testing Facility (CTF) Stage used	
Model / Type Ref.	PCA-20-1000V-PD
Additional information (if necessary may also be reported on page 2)	<input type="checkbox"/> Additional information on page 2 The risk management requirements of the standard were not addressed
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005/AMD1:2012 Comments: inclusive National differences for USA, Canada
As shown in the Test Report Ref. No. which forms part of this Certificate	2046-2-2019

This CB Test Certificate is issued by the National Certification Body

Ukrainian Scientific and Technical Institute for Certification and Testing of Electrical
Equipment UkrTEST of SE "Ukrmetrteststandart"
4, Metrologichna, str., UA-03143 Kyiv, Ukraine



IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product Capacitor charging power supply

Name and address of the applicant OOO "OEM Tech"
Odoevskogo 129-106, 220018 Minsk, Belarus

Name and address of the manufacturer OOO "OEM Tech"
Odoevskogo 129-106, 220018 Minsk, Belarus

Name and address of the factory Additional information on page 2
OOO "OEM Tech"
Odoevskogo 129-106, 220018 Minsk, Belarus
Note: When more than one factory, please report on page 2

Ratings and principal characteristics Input:110-240VAC, 50-60Hz, 20,5-9,5 A, Class I
Output: 1000VDC

Trademark (if any) OEM Tech

Customer's Testing Facility (CTF) Stage used

Model / Type Ref. PCA-20-1000V-PD


Additional information (if necessary may also be reported on page 2) Additional information on page 2

A sample of the product was tested and found to be in conformity with IEC 60601-1-2:2014
Comments:
inclusive National differences for Canada

As shown in the Test Report Ref. No. which forms part of this Certificate 2045-5-2019

This CB Test Certificate is issued by the National Certification Body

Ukrainian Scientific and Technical Institute for Certification and Testing of Electrical Equipment UkrTEST of SE "Ukrmetrteststandart"
4, Metrologichna, str., UA-03143 Kyiv, Ukraine





CERTIFICATE

CERTIFICATE OF CONFORMITY

With EN 60601-1: 2006 + A11:2011 + A1:2013, IEC 60601-1: 2005 + CORR. 1:2006 + CORR. 2:2007 + AM1:2012 (or IEC 60601-1:2012 reprint); ANSI/AAMI ES60601-1:2005/(R):2012 +A1:2012+ C1:2009+ A2:2010 standards

Certificate Number OEMSAF_EN.34518C

This certificate of conformity has been granted to the applicant based on the results of tests and evaluations, performed by Hermon Laboratories 23 October – 27 November, 2019 on representative samples of the specified products.

Product description

Tested item: Capacitor charging power supply
Model: PCA-25-350V-PD
Serial number: 0834
Hardware version: 1.4.XX
Software release: N/A

Applicant / Manufacturer's details

Name: OEM Tech.
Address: Odoevskogo 129, Minsk, 220018, Belarus

Measurement/test results are contained in the test report: OEMSAF_EN.34518.
The comments in the associated test report shall be taken into account and used in conjunction with this certificate

Michael Brun,
Product Safety Group Manager
Hermon Laboratories Ltd.

January 16, 2020

Page 1 of 1

EXPERTS IN GLOBAL COMPLIANCE SOLUTIONS



EMC



Radio



Telecom



Environmental



Product Safety



International Approvals

Hermon Laboratories Ltd.

Hatachana St., POB 23, Binyamina 30500 Israel

Phone: +972 4 628 8001, Fax: +972 4 628 8277

Email: mail@hermonlabs.com, www.hermonlabs.com



OEM Tech

LASER ELECTRONICS

OEM Tech Sp. z o.o. declares under its sole responsibility that the product(s) to which this declaration relates, is in conformity with the following directives, standard(s) and other normative document(s).

List of models:

PCA-20-500V-PD, PCA-20-800V-PD
PCA-20-1000V-PD, PCA-25-350V-PD

Directives:

Medical Device Directive
93/42/EEC
Electromagnetic Compatibility Directive
2014/30/EU
Restriction of Hazardous Substances Directive
2011/65/EU

Product Safety Standards:

IEC 60601-1:2005/AMD1:2012

EMC Standards:

IEC 60601-1-2:2014

Note: This product is classified as component type for building-in use. The EMC characteristic and safety approval of the final application has to be performed by the end-product manufacturer in accordance to the applicable standards for the end-product.

OEM Tech Sp. z o.o.
NIP: 701 041561 REGON: 149310124
KRS: 0000900002



May 05, 2025

/ Ruslan Kulbitski, Deputy Director /

OEM Tech Sp. z o.o.
Marszałkowska st. 58
00-545 Warsaw
Poland

T. +48 733 575 202
info@oem-tech.pl
www.oem-tech.pl/
VAT: PL7011041561

